



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION
Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 1205T

**Tube with Pyramid Receiver,
Ti**

Weight: 128 grams
Weight Limit: 170kg / 375lbs

Torque recommendation:
Alignment adjustment set screw:15Nm



1205T IFU REV. E (2-5-25)




stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION
Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 1205T

**Tube with Pyramid Receiver,
Ti**

Weight: 128 grams
Weight Limit: 170kg / 375lbs

Torque recommendation:
Alignment adjustment set screw:15Nm



1205T IFU REV. E (2-5-25)




stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION
Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 1205T

**Tube with Pyramid Receiver,
Ti**

Weight: 128 grams
Weight Limit: 170kg / 375lbs

Torque recommendation:
Alignment adjustment set screw:15Nm



1205T IFU REV. E (2-5-25)




stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION
Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 1205T

**Tube with Pyramid Receiver,
Ti**

Weight: 128 grams
Weight Limit: 170kg / 375lbs

Torque recommendation:
Alignment adjustment set screw:15Nm



1205T IFU REV. E (2-5-25)

1. Application
The 1205T, Tube with Pyramid Receiver, Ti is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The Tube with Receiver, Ti is used to connect any pyramid system and 30mm tube clamp. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply locite to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

4. Environment:
Store at room temperature, and avoid extremely hot, cold, or moist storage environments Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]

Do not store in wet or very hot environment



2

5. Declaration of Conformity
ST&G as manufacturer declares that the 1205T, Tube with Receiver, conforms to the requirements of Council EU 2017/745 (EU). ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any pyramid system and 30mm tube clamp. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply locite to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

4. Environment:
Store at room temperature, and avoid extremely hot, cold, or moist storage environments Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]

Do not store in wet or very hot environment



3

1. Application
The 1205T, Tube with Pyramid Receiver, Ti is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The Tube with Receiver, Ti is used to connect any pyramid system and 30mm tube clamp. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply locite to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

4. Environment:
Store at room temperature, and avoid extremely hot, cold, or moist storage environments Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]

Do not store in wet or very hot environment



2

1. Application
The 1205T, Tube with Pyramid Receiver, Ti is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The Tube with Receiver, Ti is used to connect any pyramid system and 30mm tube clamp. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply locite to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

4. Environment:
Store at room temperature, and avoid extremely hot, cold, or moist storage environments Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]

Do not store in wet or very hot environment



2

1. Application
The 1205T, Tube with Pyramid Receiver, Ti is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The Tube with Receiver, Ti is used to connect any pyramid system and 30mm tube clamp. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply locite to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

4. Environment:
Store at room temperature, and avoid extremely hot, cold, or moist storage environments Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]

Do not store in wet or very hot environment



2

5. Declaration of Conformity
ST&G as manufacturer declares that the 1205T, Tube with Receiver, conforms to the requirements of Council EU 2017/745 (EU). ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any pyramid system and 30mm tube clamp. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply locite to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

4. Environment:
Store at room temperature, and avoid extremely hot, cold, or moist storage environments Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]

Do not store in wet or very hot environment



3

1. Application
The 1205T, Tube with Pyramid Receiver, Ti is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The Tube with Receiver, Ti is used to connect any pyramid system and 30mm tube clamp. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply locite to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

4. Environment:
Store at room temperature, and avoid extremely hot, cold, or moist storage environments Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]

Do not store in wet or very hot environment



2



Notice
Tighten all the screws, which are side by side, using the torque wrench

3