



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3212T

HD Tube Clamp for 375 lbs

Weight: 99 grams
Weight Limit: 170kg / 375lbs
Recommended Torque:
Alignment adjustment set screw: 15 Nm
Pylon clamp screw: 7 Nm



3212T IFU REV. D (07-12-23)



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3212T

HD Tube Clamp for 375 lbs

Weight: 99 grams
Weight Limit: 170kg / 375lbs
Recommended Torque:
Alignment adjustment set screw: 15 Nm
Pylon clamp screw: 7 Nm



3212T IFU REV. D (07-12-23)



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3212T

HD Tube Clamp for 375 lbs

Weight: 99 grams
Weight Limit: 170kg / 375lbs
Recommended Torque:
Alignment adjustment set screw: 15 Nm
Pylon clamp screw: 7 Nm



3212T IFU REV. D (07-12-23)



stngco.com


CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3212T

HD Tube Clamp for 375 lbs

Weight: 99 grams
Weight Limit: 170kg / 375lbs
Recommended Torque:
Alignment adjustment set screw: 15 Nm
Pylon clamp screw: 7 Nm



3212T IFU REV. D (07-12-23)

1. Application
The 3212T, HD Tube Clamp for 375 lb, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The 3212T, HD Tube Clamp for 375 lb, TI is used to connect any component with 30 mm tube system. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply recommended torque to pylon clamp screws and alignment adjustment screws for connection. Apply locite to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

4. Declaration of Conformity
ST&G as manufacturer declares that the 3212T, HD Tube Clamp for 375 lb, TI conforms to the requirements of Council EU 2017/745 (MDR).

ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

Notice



Tighten all the screws, which are side by side, using the torque wrench

3

1. Application
The 3212T, HD Tube Clamp for 375 lb, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The 3212T, HD Tube Clamp for 375 lb, TI is used to connect any component with 30 mm tube system. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply recommended torque to pylon clamp screws and alignment adjustment screws for connection. Apply locite to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

2

4. Declaration of Conformity
ST&G as manufacturer declares that the 3212T, HD Tube Clamp for 375 lb, TI conforms to the requirements of Council EU 2017/745 (MDR).

ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

Notice



Tighten all the screws, which are side by side, using the torque wrench

3

1. Application
The 3212T, HD Tube Clamp for 375 lb, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The 3212T, HD Tube Clamp for 375 lb, TI is used to connect any component with 30 mm tube system. The four set screws are used for anchoring the mating pyramid system as well as for making angular adjustment for alignment. Apply recommended torque to pylon clamp screws and alignment adjustment screws for connection. Apply locite to all screws once the alignment and connection is finalized.

3. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 years warranty.

2

4. Declaration of Conformity
ST&G as manufacturer declares that the 3212T, HD Tube Clamp for 375 lb, TI conforms to the requirements of Council EU 2017/745 (MDR).

ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

Notice



Tighten all the screws, which are side by side, using the torque wrench

3