



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113




ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3230T

**HD 4 Ear Connector for
375lbs, Ti**

Weight: 76 grams
Weight Limit: 170kg / 375lbs



3230T IFU REV. D (07-12-23)



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3230T

**HD 4 Ear Connector for
375lbs, Ti**

Weight: 76 grams
Weight Limit: 170kg / 375lbs



3230T IFU REV. D (07-12-23)



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113




ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3230T

**HD 4 Ear Connector for
375lbs, Ti**

Weight: 76 grams
Weight Limit: 170kg / 375lbs



3230T IFU REV. D (07-12-23)



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 3230T

**HD 4 Ear Connector for
375lbs, Ti**

Weight: 76 grams
Weight Limit: 170kg / 375lbs



3230T IFU REV. D (07-12-23)

1. Application
The 3230T HD 4 Ear Connector for 357 1b, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The 3230T HD 4 Ear Connector for 375 1b, TI is integrated into the prosthetic socket's laminate as lamination anchor. The enclosed lamination dummy covers the pyramid during the lamination process and makes it possible to accurately fit the laminated brim. For reinforcement, the off the individual layers of stockinette between the lamination dummy and lamination anchor.

3. Reinforcement recommendations
The type of reinforcement used depends on the type of loading expected. It is up to the clinician to determine adequate layup and reinforcement for the given patient usage and activity level.

4. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 year warranty

2

5. Declaration of Conformity
ST&G as manufacturer declares that the 3230T HD 4 Ear Connector for 375 1b, TI conforms to the requirements of Council EU 2017/745 (MDR). ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.



Tighten all the screws, which are side by side, using the torque wrench

3

5. Declaration of Conformity

ST&G as manufacturer declares that the 3230T HD 4 Ear Connector for 375 1b, TI conforms to the requirements of Council EU 2017/745 (MDR). ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.



Tighten all the screws, which are side by side, using the torque wrench

3. Reinforcement recommendations
The type of reinforcement used depends on the type of loading expected. It is up to the clinician to determine adequate layup and reinforcement for the given patient usage and activity level.

4. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 year warranty

2

1. Application
The 3230T HD 4 Ear Connector for 357 1b, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations.

2. Description
The 3230T HD 4 Ear Connector for 375 1b, TI is integrated into the prosthetic socket's laminate as lamination anchor. The enclosed lamination dummy covers the pyramid during the lamination process and makes it possible to accurately fit the laminated brim. For reinforcement, the off the individual layers of stockinette between the lamination dummy and lamination anchor.

3. Reinforcement recommendations
The type of reinforcement used depends on the type of loading expected. It is up to the clinician to determine adequate layup and reinforcement for the given patient usage and activity level.

4. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 year warranty

2

1. Application

The 3230T HD 4 Ear Connector for 357 1b, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations.
2. Description
The 3230T HD 4 Ear Connector for 375 1b, TI is integrated into the prosthetic socket's laminate as lamination anchor. The enclosed lamination dummy covers the pyramid during the lamination process and makes it possible to accurately fit the laminated brim. For reinforcement, the off the individual layers of stockinette between the lamination dummy and lamination anchor.

3. Reinforcement recommendations

The type of reinforcement used depends on the type of loading expected. It is up to the clinician to determine adequate layup and reinforcement for the given patient usage and activity level.

4. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 year warranty

2

5. Declaration of Conformity
ST&G as manufacturer declares that the 3230T HD 4 Ear Connector for 375 1b, TI conforms to the requirements of Council EU 2017/745 (MDR). ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.



Tighten all the screws, which are side by side, using the torque wrench

3

5. Declaration of Conformity

ST&G as manufacturer declares that the 3230T HD 4 Ear Connector for 375 1b, TI conforms to the requirements of Council EU 2017/745 (MDR). ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.



Tighten all the screws, which are side by side, using the torque wrench

3