



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 5220T

**4 Ear Connector for 34mm
System, Ti**

Weight: 141 grams
Weight Limit: 226kg / 500lbs



5220T IFU REV. E (12-17-24)



stngco.com


CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 5220T

**4 Ear Connector for 34mm
System, Ti**

Weight: 141 grams
Weight Limit: 226kg / 500lbs



5220T IFU REV. E (12-17-24)



stngco.com

CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 5220T

**4 Ear Connector for 34mm
System, Ti**

Weight: 141 grams
Weight Limit: 226kg / 500lbs



5220T IFU REV. E (12-17-24)



stngco.com


CORPORATE OFFICE

2691 Saturn Street, Brea, CA 92821
Tel: (714) 524-0663, Fax: (714) 364-8113



ST&G CORPORATION

Rm. 911 43, Yangsan-ro
Yeongdeungpo-gu, Seoul, Korea



MDSS GmbH

Schiffgraben 41, 30175 Hannover, Germany



REF: 5220T

**4 Ear Connector for 34mm
System, Ti**

Weight: 141 grams
Weight Limit: 226kg / 500lbs



5220T IFU REV. E (12-17-24)

1. Application
The 5220T 4 Ear Connector for 34mm System, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations. Only as a system with approved compatible components.

2. Description
The 4 Ear Connector 34mm System TI is integrated into the prosthetic socket's lamination as lamination anchor For reinforcement, tie off the individual layers of stockinette into the lamination groove, and tape off the pyramid.

3. Reinforcement recommendations
The type of reinforcement used depends on the type of loading expected. It is up to the clinician to determine adequate layup and reinforcement for the given patient usage and activity level.

4. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 year warranty.

1. Application
The 5220T 4 Ear Connector for 34mm System, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations. Only as a system with approved compatible components.

2. Description
The 4 Ear Connector 34mm System TI is integrated into the prosthetic socket's lamination as lamination anchor For reinforcement, tie off the individual layers of stockinette into the lamination groove, and tape off the pyramid.

3. Reinforcement recommendations
The type of reinforcement used depends on the type of loading expected. It is up to the clinician to determine adequate layup and reinforcement for the given patient usage and activity level.

4. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 year warranty.

2

2

5. Environment:
Store at room temperature, and avoid extremely hot, cold, or moist storage environments
Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]
Do not store in wet or very hot environment

6. Declaration of Conformity
ST&G as manufacturer declares that the 5220T, 4 Ear Connector 34mm System, TI conforms to the requirements of Council EU 2017/745 (EU).

ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.

4. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 year warranty.



3

3

1. Application
The 5220T 4 Ear Connector for 34mm System, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations. Only as a system with approved compatible components.

2. Description
The 4 Ear Connector 34mm System TI is integrated into the prosthetic socket's lamination as lamination anchor For reinforcement, tie off the individual layers of stockinette into the lamination groove, and tape off the pyramid.

3. Reinforcement recommendations
The type of reinforcement used depends on the type of loading expected. It is up to the clinician to determine adequate layup and reinforcement for the given patient usage and activity level.

4. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 year warranty.

2

2

5. Environment:
Store at room temperature, and avoid extremely hot, cold, or moist storage environments
Storage Temperature Range: between temperatures of -20°C to 80°C [-4°F and 176°F]
Do not store in wet or very hot environment

6. Declaration of Conformity
ST&G as manufacturer declares that the 5220T, 4 Ear Connector 34mm System, TI conforms to the requirements of Council EU 2017/745 (EU).

ST&G products are manufactured and tested for a particular weight and activity impact level. Use by another user for whom it was not originally manufactured may cause injury and voids any written or implied warranty. The liability shall fall upon those that resale/provide the product to anyone other than the original user delivered to. Not to be resold by private party.



3

3

1. Application
The 5220T 4 Ear Connector for 34mm System, TI is to be exclusively used for the prosthetic fitting of lower extremity amputations. Only as a system with approved compatible components.

2. Description
The 4 Ear Connector 34mm System TI is integrated into the prosthetic socket's lamination as lamination anchor For reinforcement, tie off the individual layers of stockinette into the lamination groove, and tape off the pyramid.

3. Reinforcement recommendations
The type of reinforcement used depends on the type of loading expected. It is up to the clinician to determine adequate layup and reinforcement for the given patient usage and activity level.

4. Warranty
Warranty applies only when the product is used as intended without modification, following all manufacturer's recommendations. ST&G offers 2 year warranty.

